Journal of Radiography and Imaging Technology Volume-1, Issue:1- September - 2025

VISWAM PUBLICATIONS

Journal of Radiography and Imaging Technology

https://viswampublications.com/journals/index.php/jrit



Principles and Clinical applications of Photon-Counting Detector Computed Tomography (PCD CT) -A Comprehensive Review

Mr.Antonyruban S[1] , Mr.Thanga kumaran. R[2], Prof.V.Kalaivani [3], Dr.Tamijeselvan S[4]

[1] PhD Scholar, Assistant Professor, Department of Radiology and Imaging Technology, School of Allied Health Sciences, Vinayaka Mission's Research Foundation DU, Salem.
[2]Assistant Professor,Department of Radiology and Imaging Technology,School of Allied Health Sciences, Vinayaka Mission's Research Foundation-DU, Salem, Tamilnadu, India.
[3]Professor, Department of Radiology and Imaging Technology, School of Allied Health Sciences, Vinayaka Mission's Research Foundation-DU Salem, Tamilnadu, India.
[4]Asst. Professor in Radiography, Mother Theresa PG and Research Institute of Health Sciences, Puducherry

ARTICLE INFO

Received : 05.03.2025
Revised : 28.05.2025
Accepted : 11.08.2025
Published : 26.09.2025

Corresponding author:
Dr.Tamijeselvan S
tamije1970@gmail.com

Keywords:
Photon-Counting Detector CT,
Energy-Integrating Detector,
High-Resolution Imaging,
Radiation Dose Reduction,
Spectral Imaging,
Material Decomposition,
Multi-Energy CT

ABSTRACT

Photon-counting computed tomography (PCCT) represents a significant advancement in CT imaging technology. Unlike conventional energy-integrating detectors (EIDs), PCCT employs photon-counting detectors (PCDs) that directly convert X-ray photons into electrical signals. This enables enhanced spatial resolution, reduced electronic noise, and improved dose efficiency. The ability to perform multi-energy acquisition and material decomposition further extends PCCT's clinical applications. This review explores the evolution of CT detector technology, the principles of PCCT, its advantages over conventional CT, and its diverse clinical applications, including cardiovascular imaging, oncology, musculoskeletal imaging, and pediatric imaging. We also discuss challenges, limitations, and future perspectives of PCCT in clinical radiology.

This article is licensed under Creative Commons Attribution-Non-commercial 4.0 International License.
Copyright©2025 Author(s) retains the copyright of this article. DOI:

1.Introduction

Computed Tomography (CT) has revolutionized medical imaging, providing high-resolution cross-sectional images for accurate diagnosis and treatment planning. Over the years, advancements in CT technology have focused on improving image quality while minimizing radiation exposure. One of the most significant innovations in recent years is the development of **Photon-Counting Detector CT (PCD-CT)**, which represents a paradigm shift in X-ray detection and image reconstruction.

Unlike conventional **energy-integrating detectors (EID-CT)**, which measure the total energy of incoming X-rays without distinguishing photon energy levels, **photon-counting detectors (PCDs)** individually count and classify incoming X-ray photons based on their energy. This fundamental difference enables several clinical advantages, including **higher spatial resolution, improved contrast-to-noise ratio, better material decomposition, and reduced radiation dose**.

The emergence of **PCD-CT** has sparked considerable interest across various medical fields, including **cardiology, oncology, musculoskeletal imaging, and neuroimaging**, where enhanced tissue differentiation and improved diagnostic accuracy are critical. Additionally, the ability of PCD-CT to eliminate electronic noise and enhance spectral imaging capabilities has opened new avenues for personalized medicine and functional imaging applications.

This review provides a comprehensive overview of the clinical benefits of **photon-counting detector CT**, discussing its **technical principles, advantages over conventional CT, and its impact on different medical specialties**. Furthermore, we will explore current challenges and future prospects of this innovative imaging technology in clinical practice.

2.Evolution of CT Detector Technology

CT imaging has evolved through distinct technological phases:

- **Xenon Gas Detector Era (1991): Introduction of the first spiral CT scanner.**
- **Scintillator Detector Era (1998): Development of multislice CT technology.**
- **Dual-Energy CT Era (2005): Implementation of dual-source CT for spectral imaging.**
- **Photon-Counting Detector CT Era (2021): Intro duction of clinical PCCT scanners.**

JOURNEY OF CT DETECTOR



Fig: 1.0 XENON GAS CT ERA



Fig: 1.1 SCINTILLATOR CT ERA



Fig: 1.2 DUAL ENERGY CT ERA

2.1. Working Mechanism of Photon-Counting Detectors

In traditional **EID-CT**, X-ray photons pass through the patient and strike the detector, where their energy is converted into an electrical signal. However, this method does not differentiate between photons of different energy levels, leading to **lower spectral resolution and increased image noise**.




Fig 2shows a conventional energy-integrated detector (EID) CT. A scintillator converts X-rays to visible light. The light signal is received at the photodiode, which is partitioned by reflective septae to prevent optical crosstalk between detector elements.

In contrast, **PCD-CT directly detects individual photons**, measures their energy, and classifies them into different energy bins. This process enables **true multi-energy imaging**, eliminating the need for additional contrast-based spectral techniques like dual-energy CT.




Figure 3 Illustrates the detector mechanism in photon-counting detector CT. X-rays are converted to an electron cloud that moves across a high voltage within the semiconductor. There are no septae in the detector, which allows for reduced pixel size.

2.2. Key Features of PCD Technology

- **Energy Discrimination**: Differentiates photons by energy, enhancing tissue contrast.
- **Higher Spatial Resolution**: Reduces electronic noise and enhances fine-detail imaging.
- **Improved Dose Efficiency**: Low-energy photons, often discarded in conventional CT, are utilized, allowing for lower radiation doses.
- **Elimination of Electronic Noise**: Unlike EID-CT, which suffers from background noise, PCD-CT detects only true photon interactions.

3. Advantages Over Conventional CT

The transition from EID-CT to PCD-CT offers several advantages that significantly improve diagnostic capabilities:

3.1. Improved Spatial Resolution

- PCD-CT eliminates the need for scintillator-based conversion, reducing blur and **enhancing the visualization of small structures** (e.g., lung nodules, vascular plaques).

3.2. Enhanced Material Decomposition

- With true spectral imaging, PCD-CT can differentiate between materials such as **calcium, iodine, uric acid, and fat**, aiding in conditions like **gout, osteoporosis, and tumor characterization**.

3.3. Reduced Radiation Dose

- Since PCD-CT makes use of lower-energy photons, it enables **diagnostic-quality imaging at lower doses**, which is particularly beneficial for **pediatric, oncology, and follow-up imaging**.

3.4. Artifact Reduction

- Traditional CT scans suffer from **beam-hardening artifacts** due to polychromatic X-ray beams. PCD-CT mitigates this issue, improving the accuracy of **metal implant imaging and bone assessments**.

4. Clinical Impact Across Different Medical Specialties

4.1. Cardiology

- **Improved Coronary Artery Imaging**: PCD-CT enhances **plaque characterization and stent visualization**, providing more accurate assessment of coronary artery disease.




Fig 4 shows coronary CTA: Optimal image quality with low image noise and high contrast resolution allows for the reliable exclusion of coronary artery disease. Left: 3D-VRT and right: curved multiplanar reformation (cMPR) of RCA, LAD, and CX arteries show absence of coronary plaque and stenosis.

- **Lower Contrast Requirements**: Reduces contrast agent dosage, benefiting patients with kidney disease.

4.2. Oncology

- **Better Tumor Characterization**: Differentiates **tumor tissue from normal tissue** with greater precision using spectral imaging.
- **Lower Dose for Cancer Follow-Ups**: PCD-CT enables safer, repeated imaging for **radiation-sensitive patients**.

4.3. Musculoskeletal Imaging

- **Superior Bone and Joint Imaging**: High-resolution imaging enables better detection of **microfractures, osteoporosis, and joint erosions**.
- **Reduced Metal Artifacts**: Enhances post-surgical imaging in patients with **joint replacements or spinal implants**.

4.4. Neurology

- **Higher Resolution Brain Imaging**: PCD-CT improves the detection of **microbleeds, ischemic stroke, and brain tumors**.




Fig 5 shows patients with acute ischemic stroke treated using mechanical thrombectomy. (a) The original images of mobile PCD-CT, (b) iodine overlay map, (c) VNC image, and (d) follow-up non-contrast CT (NCCT) image.

- **Better Differentiation of Hemorrhages**: Energy-based photon detection enhances the visualization of **intracranial hemorrhages**.

4.5. Pulmonology

- **Early Lung Cancer Detection**: High-resolution imaging allows for the identification of **subtle pulmonary nodules**.




Fig 6 Shows Inspiratory photon-counting detector (PCD)-CT images in a 6-year-old girl with cystic fibrosis show structural lung features assessed in Brody II scoring. a, b Axial inspiratory image (a) illustrates varicose bronchiectasis (arrow) in the right middle lobe. Coronal expiratory image (b) shows air trapping (arrow) in the left lower lobe. c, d Sagittal images show mucus plugging within a dilated bronchus (arrow in c), ground-glass opacity (arrow in d) and peribronchial cuffing (arrowhead in c, d). Images are displayed in standard lung windows (width: 1,500; level: –500). The axial image is 0.8 mm and coronal and sagittal images are 3 mm

- **Better Assessment of Interstitial Lung Disease**: Improved contrast-to-noise ratio enhances lung parenchymal evaluation.

5. Challenges and Limitations of PCCT

5.1 Cost and Implementation Barriers

- Higher initial costs due to advanced detector technology.
- Need for integration with existing radiology infrastructure.

5.2 Data Processing and Storage

- Large datasets require high computational power.
- Advanced algorithms are needed for efficient image reconstruction.

5.3 Sensitivity to Artifacts and Noise

- Low-dose protocols may still be affected by photon starvation artifacts.
- Optimization of energy thresholds is required for specific clinical scenarios.

6. Future Directions and Research Opportunities

6.1 Machine Learning and AI Integration

- AI-based noise reduction and artifact correction algorithms.
- Automated material decomposition for enhanced tissue characterization.

6.2 Hybrid Imaging Approaches

- Combination of PCCT with MRI for multi-modal imaging.
- Integration with functional imaging techniques for improved disease assessment.

6.3 Expanded Clinical Studies

- Large-scale trials to validate clinical benefits in diverse patient populations.

Further research into optimal scan protocols for different diseases

7. Conclusion

Photon-Counting Detector CT represents a **major advancement in medical imaging**, offering **superior spatial resolution, improved spectral imaging, and lower radiation exposure** compared to conventional EID-CT. Its applications span across multiple medical specialties, enhancing **diagnostic accuracy, treatment planning, and patient outcomes**. As research and development continue, **PCD-CT is poised to become the future standard for CT imaging**.

References

1. The Organisation for Economic Co-operation and Development. (2018). *Computed tomography (CT) exams (indicator)*. OECD website. https://data.oecd.org/healthcare/computed-tomography-ct-exams.htm
2. Shefer, E., Altman, A., Behling, R., et al. (2013). State of the art of CT detectors and sources: A literature review. *Current Radiology Reports, 1*(1), 76–91.
3. Johnson, T., Fink, C., Schönberg, S. O., & Reiser, M. F. (Eds.). (2011). *Dual energy CT in clinical practice.* Springer.
4. McCollough, C. H., Leng, S., Yu, L., & Fletcher, J. G. (2015). Dual- and multi-energy CT: Principles, technical approaches, and clinical applications. *Radiology, 276*(3), 637–653.
5. Johnson, T. R. (2012). Dual-energy CT: General principles. *AJR American Journal of Roentgenology, 199*(5 Suppl), S3–S8.
6. Johnson, T. R., Krauss, B., Sedlmair, M., et al. (2007). Material differentiation by dual energy CT: Initial experience. *European Radiology, 17*(6), 1510–1517.
7. Pourmorteza, A., Symons, R., Sandfort, V., et al. (2016). Abdominal imaging with contrast-enhanced photon-counting CT: First human experience. *Radiology, 279*(1), 239–245.
8. Yu, Z., Leng, S., Jorgensen, S. M., et al. (2016). Evaluation of conventional imaging performance in a research whole-body CT system with a photon-counting detector array. *Physics in Medicine & Biology, 61*(4), 1572–1595. https://doi.org/10.1088/0031-9155/61/4/1572
9. Bennett, J. R., Opie, A. M., Xu, Q., et al. (2014). Hybrid spectral micro-CT: System design, implementation, and preliminary results. *IEEE Transactions on Biomedical Engineering, 61*(2), 246–253.
10. Iwanczyk, J. S., Nygård, E., Meirav, O., et al. (2009). Photon counting energy dispersive detector arrays for X-ray imaging. *IEEE Transactions on Nuclear Science, 56*(3), 535–542.
11. Schlomka, J. P., Roessl, E., Dorscheid, R., et al. (2008). Experimental feasibility of multi-energy photon-counting K-edge imaging in pre-clinical computed tomography. *Physics in Medicine & Biology, 53*(15), 4031–4047.
12. Shikhaliev, P. M. (2008). Energy-resolved computed tomography: First experimental results. *Physics in Medicine & Biology, 53*(20), 5595–5613.
13. Taguchi, K., &Iwanczyk, J. S. (2013). Vision 20/20: Single photon counting X-ray detectors in medical imaging. *Medical Physics, 40*(10), 100901.
14. Xu, C., Danielsson, M., Karlsson, S., Svensson, C., &Bornefalk, H. (2012). Preliminary evaluation of a silicon strip detector for photon-counting spectral CT. *Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment, 677*, 45–51.
15. Anderson, N. G., Butler, A. P., Scott, N. J., et al. (2010). Spectroscopic (multi-energy) CT distinguishes iodine and barium contrast material in mice. *European Radiology, 20*(9), 2126–2134.
16. Rajendran, K., Löbker, C., Schon, B. S., et al. (2017). Quantitative imaging of excised osteoarthritic cartilage using spectral CT. *European Radiology, 27*(1), 384–392.
17. Xu, Q., Yu, H., Bennett, J., et al. (2012). Image recon-

struction for hybrid true-color micro-CT. *IEEE Transactions on Biomedical Engineering, 59*(6), 1711–1719.

18. Yu, Z., Leng, S., Kappler, S., et al. (2015). A prototype whole-body PCCT system: Initial results in phantom, cadavers, and swine. *The 3rd Workshop on Medical Applications of Spectroscopic X-Ray Detectors.*

19. Gutjahr, R., Halaweish, A. F., Yu, Z., et al. (2016). Human imaging with photon counting-based computed tomography at clinical dose levels: Contrast-to-noise ratio and cadaver studies. *Investigative Radiology, 51*(7), 421–429.

20. Muenzel, D., Bar-Ness, D., Roessl, E., et al. (2017). Spwectral photon-counting CT: Initial experience with dual-contrast agent K-edge colonography. *Radiology, 283*(3), 723–729.